



**EA MLA Signatory**  
**Český institut pro akreditaci, o.p.s.**  
**Olšanská 54/3, 130 00 Praha 3**

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

## **CERTIFICATE OF ACCREDITATION**

No. **64/2024**

**M & B Calibr, spol. s r.o.**  
with registered office Krumlovská 1454/26, 664 91 Ivančice,  
Company Registration No. 43389783

for the Calibration Laboratory No. **2301**  
Calibration Laboratory

Scope of accreditation:

Calibration of meters of length, plane angle, weight, speed, hardness, roughness, force, pressure, temperature, electrical quantities, time quantities and humidity to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

**ČSN EN ISO/IEC 17025:2018**

In its activities performed within the scope and for the period of validity of this Certificate, the Conformity Assessment Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Accredited Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 21/2024 of 22. 1. 2024, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **19. 7. 2024**

Prague: 14. 2. 2024



Jan Velíšek  
Director of the Department  
of Testing and Calibration Laboratories  
Czech Accreditation Institute

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**Calibration laboratory locations:**

- |                                  |   |
|----------------------------------|---|
| 1. <b>Calibration Laboratory</b> | Krumlovská 1454/26, 664 91 Ivančice         |
| 2. <b>Calibration Laboratory</b> | Strojírenská 259/16, Zličín, 155 21 Praha 5 |

**CMC for the field of measured quantity: Length**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place		
		min.	unit	max.	unit							
1	Steel parallels	0.5 mm	to	1,000 mm			(2L + 0.2) µm	Comparative measurement using steel parallels	KP D1	1		
2*	Steel length gauges	0 m	to	2 m			60 µm	Comparative measurement using steel parallels	KP D2	1,2		
		2 m	to	5 m			180 µm					
	Steel tape measures	0 m	to	2 m			0.14 mm	Comparative measurement on a reference track				
		2 m	to	3 m			0.28 mm					
		3 m	to	5 m			0.42 mm					
		5 m	to	8 m			0.70 mm					
		8 m	to	10 m			0.98 mm					
3	Tape measures	0 m	to	10 m			0.4 mm	Comparative measurement on a reference track	KP D3	1,2		
		10 m	to	20 m			0.6 mm					
		20 m	to	50 m			1.0 mm					
		50 m	to	100 m			2.2 mm					
	Laser distance meters	0 m	to	5 m			0.2 mm					
4	Limit and end measuring rings	1 mm	to	100 mm			(2L + 0.5) µm	Direct and comparative measurement by a distance meter	KP D4	1		
		100 mm	to	500 mm			(2L + 2.4) µm					
		1 mm	to	200 mm			(4L + 1.3) µm			2		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
min.	unit	max.	unit							
	Limit snap gauges	1 mm	to	100 mm			(2·L + 0.5) µm (2·L + 2.4) µm			1,2
	Feeler gauges	100 mm	to	500 mm						
	Limit cylindrical gauges	0.02 mm	to	100 mm			(2·L + 0.5) µm (2·L + 2.4) µm			
		100 mm	to	500 mm						
5*	Feeler gauges, Limit cylindrical gauges.	1 mm	to	125 mm			(2·L + 2.4) µm	Direct measurement by a micropasameter	KP D4	1
6	Limit plug gauges	1 mm	to	200 mm			(3·L + 3) µm	Direct measurement by a distance meter	KP D5	1,2
		1 mm	to	160 mm			(1·L + 4) µm	Direct measurement on MasterScanner XP 16060		1
	Threaded rings	1 mm	to	3 mm			(3·L + 3) µm	Comparison by a wear pin gauge		1
		2.5 mm	to	200 mm			(3·L + 3) µm	Comparative measurement by a distance meter		
		3 mm	to	160 mm			(1·L + 4) µm	Direct measurement on MasterScanner XP 16060		
7*	Limit plug gauges	1 mm	to	125 mm			(3·L + 3.5) µm	Direct measurement by a micropasameter	KP D5	1
8*	Slide gauges: slide rules, depth gauges, height gauges	0 mm	to	1,000 mm			12 µm 20 µm	Comparative measurement using steel parallels	KP D6	1,2
9*	Micrometer gauges: micrometers, pasameters, micropasameters, micrometer heads, micrometer depth gauges	1,000 mm	to	3,000 mm				Comparative measurement using steel parallels	KP D7	1,2
		0 mm	to	25 mm			0.7 µm 1.4 µm			
		0 mm	to	100 mm						

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
		100 mm	to	1,000 mm			2.5 µm 4.1 µm			
10*	Inside micrometers Three contact internal gauges	1,000 mm	to	1,500 mm			2.0 µm 4.0 µm	Comparative measurement using setting rings	KP D8	1,2
		2 mm	to	100 mm			4.0 µm			1
		100 mm	to	300 mm			4.0 µm			2
11	Inside micrometer gauges	100 mm	to	200 mm			(3·L + 2.2) µm	Direct measurement by a distance meter	KP D9	1
		10 mm	to	3,000 mm			(3·L + 2.2) µm			2
12*	Electromagnetic, ultrasonic thickness gauges	0 mm	to	1.5 mm			(1·L + 1.3) µm (1·L + 2.3) µm	Comparative measurement by a thickness reference standard	KP D10	1
13	Direct and lever dial indicators	1.5 mm	to	500 mm			0.3 µm	Direct measurement by a special measuring device	KP D11	1,2
	Two-contact internal gauges	0 mm	to	100 mm			0.3 µm			2
		2 mm	to	205 mm			(3·L + 2.2) µm	Direct measurement by a length gauge		
14	Gauges, jigs, templates, meters of plane and angle	205 mm	to	1 000 mm			(4.5·L + 1.7) µm	Measurement by a 3D CMM	KP D12	1
15*	Profile projectors, measuring microscopes	0 mm	to	2,000 mm			(1·L + 2.6) µm	Comparative measurement using a rule	KP D13	1
16*	Measurement of straightness, linear sensing, measurement of flatness of engineering gauges	0 mm	to	300 mm			(1·L + 0.1) µm 1.5 µm/m <sup>2</sup>	Direct measurement by a laser interferometer	KP D14	1
		0 m	to	20 m						
		0 m	to	20 m						



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
min.	unit	max.	unit					
17	Gauges, jigs, templates, meters	0 mm	to	600 mm	(2.5L + 1.2) µm (0.8L + 0.5) µm	Measurement by a linear height gauge	KP D15	1
18*	Linear height gauges	0 mm	to	600 mm	(1L + 3.0) µm	Comparative measurement by a calibration comb	KP D16	1
		600 mm	to	1,000 mm	(1L + 2.6) µm	Comparative measurement by a calibration comb and steel parallels		
19*	Contourographs	0 mm	to	100 mm	(2L + 0.2) µm	Comparative measurement using end standards	KP D17	1
20*	Length measuring instruments	0 mm	to	1,000 mm	(2L + 0.2) µm	Direct measurement by a laser interferometer	KP D18	1
21*	3D coordinate measuring machines	0 mm	to	600 mm	(2L + 0.2) µm	Comparative measurement by a calibration comb	KP D19	1
		600 mm	to	1,000 mm	(1L + 0.1) µm	Comparative measurement by a calibration comb and steel parallels		
		0 mm	to	10,000 mm	(2L + 3.5) µm	Direct measurement by a laser interferometer		
22	Gauges, meters, jigs, templates, rules	0 mm	to	330 mm	(2L + 3.5) µm (5L + 2) µm	Direct measurement by a 2D microscope	KP D20	1 2
		0 mm	to	300 mm				
23	Blade and surface rules	0 mm	to	2,000 mm	(5L + 12) µm (5L + 12) µm (5L + 12) µm (2.5L + 1.2) µm	Direct measurement on a plate Measurement on a bed Direct measurement on a plate	KP D21	1
		2,000 mm	to	3,000 mm				
		0 mm	to	1,000 mm				
		1,000 mm	to	1,500 mm				
24*	Roughness meters	0.01 µm	to	6,000 µm	5 %	Comparative measurement by a roughness reference standard	KP DR1	1
25	Roughness standards	0.01 µm	to	6,000 µm	5 %	Direct measurement by a roughness meter	KP DR1	1



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
26	Angles	0 °	to	180 °		Length up to 3 m	(4.5L + 2) µm	Direct measurement by a 3D CMM	KP R2	1
						Length up to 0.6 m	(20L + 2) µm	Direct measurement on special equipment		2

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanatory notes:

CMM – coordinate measuring machine

L – nominal length in metres



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity: Plane angle**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max	unit					
1	Levels – builder's, liquid, machine	-52 mm/m	to	52 mm/m	.	Division sensitivity from 0.01 mm/m	0.005mm/m	Direct measurement by a small angle generator	KP R1	1,2
	Clinometers	-180 °	to	180 °	.	Division from 0.01°	0.15°			
2	Angle gauges	0 °	to	360 °	.		5'	Direct measurement using angle gauges	KP R2	1,2

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity: Mass**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit					
1*	Non-automatic weighing instruments	0.001 g	to	2000 g	2.7·10 <sup>-6</sup>	Comparative measurement using a reference weight	KP VA1	1
		2 kg	to	20 kg	1.4·10 <sup>-5</sup>	E2 class weight		
		20 kg	to	1000 kg	5.0·10 <sup>-5</sup>	F2 class weight		
2	Weights and other objects	1 g			0.4 mg	Comparison with a class F2 reference weight	KP VA2	1
		2 g			0.5 mg			
		5 g			0.6 mg			
		10 g			0.7 mg			
		20 g			0.9 mg			
		50 g			1.2 mg			
		100 g			1.9 mg			
		200 g			2.8 mg			
		500 g			5.1mg			
		1 kg			10 mg			
		2 kg			17 mg			
		5 kg			31 mg			
		10 kg			60 mg			
		20 kg			90 mg			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanatory notes: The lowest expanded measurement uncertainty is stated without accounting for the effect of the calibrated meter.

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity:**

**Rotational speed**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit					
1*	Revolution meters	30 min <sup>-1</sup>	to	40000 min <sup>-1</sup>	(1.1 % + 0.5d)	Direct measurement by a revolution generator	KP OT1	1

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanatory notes:

d - division of the scale of the calibrated gauge



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity:** Hardness

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min	unit.	max	unit.					
1	Rockwell hardness plates and samples	70 HRA	to	85 HRA			0.40 HRA	Direct measurement	KP TV1	1
		60 HRB	to	100 HRB			0.40 HRB			
		20 HRC	to	70 HRC			0.40 HRC			
	Shore A hardness plates	0 ShA	to	100 ShA			2.0 ShA			
	Shore D hardness plates	0 ShD	to	100 ShD			2.0 ShD			
2*	Brinell hardness plates	8 HBW	to	650 HBW			1.0 %	Direct measurement using reference hardness plates		
	Vickers hardness plates	10 HV	to	2900 HV		HV2 to HV50	1.0 %			
	Rockwell hardness meters for metals	70 HRA	to	85 HRA			0.50 HRA			
		60 HRB	to	100 HRB			0.50 HRB			
		20 HRC	to	70 HRC			0.50 HRC			
2*	Vickers hardness meters for metals	10 HV	to	2000 HV			0.50 %			
	Brinell hardness meters for metals	10 HBW	to	650 HBW			0.50 %			
		1 Sh	to	100 Sh			0.50 Sh			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity: Force, mechanical tests**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min	unit.					
1*	Torque wrenches	0,1 Nm	to	1100 Nm	0,65 %	Comparative measurement by a reference torque sensor	KP S1	1
		1100 Nm	to	3000 Nm	0,90 %			
2	Force meters and extensometric sensors	0,1 Nm	až	500 Nm	0,40 %	Comparative measurement by a reference force sensor	KP S2	1
3*	Force meters and extensometric sensors	0 N	to	5 kN	0,20 %	Comparative measurement by a reference force sensor		
		5 kN	to	30 kN	0,30 %			
		0 N	to	5 kN	0,20 %			
		5 kN	to	20 kN	0,30 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity: Pressure**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work -place		
		min.	unit	max	unit							
1*	Deformation manometers, tyre pressure gauges, electromechanical pressure gauges (digital pressure gauges, pressure transducers with digital output of the measured quantity)	-100 kPa	to	0 kPa		Gas	Overpressure/ underpressure	130 Pa 18 Pa 130 Pa 0.1%	Comparative measurement by a pressure reference standard	KP T1, KP T2		
		0 kPa	to	35 kPa								
		35 kPa	to	160 kPa								
		160 kPa	to	2000 kPa								
		25 kPa	to	600 kPa		Liquids	Overpressure	180 Pa 0.03 % 0.05 % 0.1 % 0.2 %				
		0.6 MPa	to	6 MPa								
		6 MPa	to	60 MPa								
		60 MPa	to	70 MPa								
		70 MPa	to	140 MPa								

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity: Temperature**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place	
		min.	unit						
1*	Direct-indicating thermometers	-30 °C	to	0 °C	0.06 °C	Comparison with a reference digital thermometer in a dry block calibrator	KP TE1	1	
		0 °C	to	100 °C	0.05 °C	Comparison with a reference digital thermometer in a liquid bath			
		100 °C	to	200 °C	0.06 °C	Comparison with a reference digital thermometer in a dry block calibrator			
		200 °C	to	300 °C	0.09 °C				
		300 °C	to	400 °C	0.4 °C				
		400 °C	to	500 °C	0.5 °C				
		500 °C	to	650 °C	0.6 °C				
	Contactless thermometers	650 °C	to	1,100 °C	1.5 °C	Comparison with a reference digital thermometer in an air oven			
2*		-10 °C	to	200 °C	3.0 °C	Comparison with a reference pyrometer on target-type or cavity-type black body			
		200 °C	to	500 °C	6.0 °C				
		500 °C	to	800 °C	10.0 °C				
Thermoelectric temperature sensors	-30 °C	to	0 °C	0.7 °C	Comparison with a reference digital thermometer in a dry block calibrator	KP TE2	1		
	0 °C	to	100 °C	0.7 °C	Comparison with a reference digital thermometer in a liquid bath				
			100 °C	to	550 °C			0.9 °C	Comparison with a reference digital thermometer in a dry block calibrator
			550 °C	to	800 °C			2.3 °C	Comparison with a reference digital thermometer in an air oven



11\_01-P508b\_K-20221122

The Appendix is an integral part of  
Certificate of Accreditation No. 64/2024 of 14/02/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit					
3*	Resistance temperature sensors	-30 °C	to	0 °C	0.15 °C	Comparison with a reference digital thermometer in an oven	KP TE3	1
		0 °C	to	100 °C	0.13 °C	Comparison with a reference digital thermometer in a liquid bath		
		100 °C	to	400 °C	0.45 °C	Comparison with a reference digital thermometer in an oven		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity: Electrical quantities**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
1	DC voltage sources	0 mV	to	100 mV			0,0062 % + 6,1 µV	Direct measurement by a standard multimeter	KP EL2	1
		0,1 V	to	1 V			0,0047 % + 16 µV			
		1 V	to	10 V			0,0047 % + 0,14 mV			
		10 V	to	100 V			0,0079 % + 2,0 mV			
		100 V	to	1000 V			0,0079 % + 20 mV			
	DC voltage meters	0 mV	to	200 mV			0,0053 % + 7,7 µV	Direct generation with a standard calibrator	KP EL1	
		0,2 V	to	2 V			0,0028 % + 15 µV			
		2 V	to	20 V			0,0028 % + 0,15 mV			
		20 V	to	200 V			0,0028 % + 1,5 mV			
		200 V	to	1000 V			0,0035 % + 17 mV			
2	Direct current sources	0 µA	to	10 µA			0,050 % + 6,1 nA	Direct measurement by a standard multimeter	KP EL2	1
		10 µA	to	100 µA			0,074 % + 17 nA			
		0,1 mA	to	1 mA			0,075 % + 0,16 µA			
		1 mA	to	10 mA			0,034 % + 1,2 µA			
		10 mA	to	100 mA			0,034 % + 12 µA			
		100 mA	to	1 A			0,068 % + 0,15 mA			
		1 A	to	3 A			0,061 % + 1,1 mA			
	Direct current meters	0 µA	to	200 µA			0,020 % + 73 nA	Direct generation with a standard calibrator	KP EL1	
		0,2 mA	to	2 mA			0,016 % + 0,13 µA			
		2 mA	to	20 mA			0,009 % + 0,94 µA			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
		2 A	to	30 A			0,069 % + 2,5 mA			
3	AC voltage sources	0 A	to	1500 A			0,42 % + 0,13 A	Simulation using current coil		
		0,1 mV	to	100 mV		10 Hz to 20 kHz	0,086 % + 36 µV	Direct measurement by a standard multimeter	KP EL2	1
						20 kHz to 50 kHz	0,16 % + 59 µV			
						50 kHz to 100 kHz	0,69 % + 93 µV			
		0,1 V	to	1 V		10 Hz to 20 kHz	0,076 % + 0,43 mV			
						20 kHz to 50 kHz	0,14 % + 0,73 mV			
						50 kHz to 100 kHz	0,71 % + 0,81 mV			
		1 V	to	10 V		10 Hz to 20 kHz	0,076 % + 4,4 mV			
						20 kHz to 50 kHz	0,14 % + 7,3 mV			
						50 kHz to 100 kHz	0,71 % + 17 mV			
		10 V	to	100 V		10 Hz to 20 kHz	0,076 % + 44 mV			
						20 kHz to 50 kHz	0,14 % + 73 mV			
						50 kHz to 100 kHz	0,71 % + 81 mV			
		100 V	to	750 V		10 Hz to 20 kHz	0,07 % + 0,35 V			
						20 kHz to 50 kHz	0,14 % + 0,59 V			
						50 kHz to 100 kHz	0,61 % + 1,5 V			
	AC voltage meters	0,1 mV	to	200 mV		10 Hz to 45 Hz	0,16 % + 64 µV	Direct generation with a standard calibrator	KP EL1	
						45 Hz to 10 kHz	0,049 % + 69 µV			
		0,2 V	to	2 V		10 kHz to 20 kHz	0,12 % + 86 µV			
						20 kHz to 50 kHz	0,20 % + 98 µV			
						10 Hz to 45 Hz	0,082 % + 0,38 mV			
						45 Hz to 10 kHz	0,042 % + 0,29 mV			
						10 kHz to 20 kHz	0,12 % + 0,46 mV			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
		2 V	to	20 V		20 kHz to 50 kHz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz	0,17 % + 0,73 mV 0,084 % + 3,8 mV 0,042 % + 2,7 mV 0,12 % + 4,7 mV 0,17 % + 5,6 mV			
		20 V	to	200 V		30 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 40 kHz	0,082 % + 31 mV 0,10 % + 43 mV 0,14 % + 56 mV 0,17 % + 61 mV			
		200 V	to	1000 V		30 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0,082 % + 0,45 V 0,09 % + 0,28 V 0,14 % + 0,38 V 0,16 % + 0,49 V			
4	Alternating current sources	0,1 µA	to	100 µA		10 Hz to 1 kHz	0,15 % + 7,3 nA	Direct measurement by a standard multimeter	KP EL2	1
		0,1 mA	to	1 mA		10 Hz to 5 kHz	0,10 % + 0,51 µA			
		1 mA	to	10 mA		10 Hz to 5 kHz	0,093 % + 5,1 µA			
		10 mA	to	100 mA		10 Hz to 5 kHz	0,093 % + 51 µA			
		0,1 A	to	1 A		10 Hz to 5 kHz	0,095 % + 0,51 mA			
		1 A	to	3 A		10 Hz to 5 kHz	0,11 % + 3,4 mA			
	Alternating current meters	20 µA	to	200 µA		10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz 10 Hz to 45 Hz 45 Hz to 1 kHz	0,12 % + 71 nA 0,02 % + 40 nA 0,17 % + 0,10 µA 0,35 % + 0,18 µA 0,13 % + 0,15 µA 0,019 % + 90 nA	Direct generation with a standard calibrator	KP EL1	
		0,2 mA	to	2 mA						

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
		2 mA	to	20 mA		1 kHz to 10 kHz 10 kHz to 30 kHz 10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0,087 % + 0,33 µA 0,20 % + 0,57 µA 0,057 % + 1,6 µA 0,048 % + 0,82 µA 0,064 % + 1,8 µA 0,11 % + 3,5 µA			
		20 mA	to	200 mA		10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 30 kHz	0,13 % + 16 µA 0,019 % + 8,0 µA 0,087 % + 35 µA 0,14 % + 39 µA			
		0,2 A	to	2 A		10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 30 kHz	0,13 % + 0,16 mA 0,024 % + 82 µA 0,088 % + 0,35 mA 0,26 % + 2,1 mA 0,50 % + 3,0 mA			
		2 A	to	30 A		30 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0,053 % + 1,6 mA 0,042 % + 0,89 mA 0,10 % + 3,0 mA 0,12 % + 3,5 mA 0,64 % + 15 mA			
		0 A	to	1500 A		30 Hz to 60 Hz	0,42 % + 0,13 A	Simulation using current coil		
5	DC resistance / DC resistance meters			0,1 Ω 1 Ω 10 Ω 100 Ω			8,7 mΩ 8,9 mΩ 11 mΩ 20 mΩ	Direct generation with a standard calibrator	KP EL 1	1

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
				1 kΩ			0,16 Ω			
				10 kΩ			1,6 Ω			
				100 kΩ			15 Ω			
				1 MΩ			0,29 kΩ			
				10 MΩ			8,0 kΩ			
				100 MΩ			0,98 MΩ			
				1 GΩ			22 MΩ			
		0 Ω	to	100 Ω			0,018 % + 88 m Ω			
		100 Ω	to	330 Ω			0,012 % + 0,11 Ω			
		330 Ω	to	1 kΩ			0,01 % + 0,19 Ω			
		1 kΩ	to	3,3 kΩ			0,012 % + 0,29 Ω			
		3,3 kΩ	to	10 kΩ			0,0082 % + 1,4 Ω			
		10 kΩ	to	33 kΩ			0,012 % + 2,2 Ω			
		33 kΩ	to	100 kΩ			0,0066 % + 18 Ω			
		100 kΩ	to	330 kΩ			0,011 % + 24 Ω			
		330 kΩ	to	1 MΩ			0,0066 % + 0,18 kΩ			
		1 MΩ	to	3,3 MΩ			0,013 % + 0,24 kΩ			
		3,3 MΩ	to	10 MΩ			0,0075 % + 1,8 kΩ			
		10 MΩ	to	33 MΩ			0,052 % + 6,4 kΩ			
		33 MΩ	to	100 MΩ			0,064 % + 0,21 MΩ			
		100 MΩ	to	330 MΩ			1,3 % + 1,9 MΩ			
		330 MΩ	to	1 GΩ			2,4 % + 13 MΩ			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**M & B Calibr, spol. s r.o.**  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

**CMC for the field of measured quantity: Time quantities**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit					
1	Time interval / time meters, stopwatches, timers	1 s	to	86400 s	0,5 s	Comparison with a reference digital stopwatch	KP Č1	1

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.  
CAB number 2301, Calibration Laboratory  
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Humidity

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work -place
		min.	unit					
1*	Relative humidity meters except psychrometers	10 % RH	to	95 % RH	(20 to 40) °C	2.3% RH	Comparative measurement by a reference hygrometer	KP VL1 1

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

